



From:
Trace Analytics, LLC
15768 Hamilton Pool Road
Austin, Texas 78738

800-247-1024 • 512-263-0000
Fax 512-263-0002
E-mail service@AirCheckLab.com

Report 17-18938, Sampled on
6/21/2017

Next Sample Due Quarterly, Approximately
9/21/2017

To:
Total Diving
6356 Sherbrooke West
Montreal, QC H4B 1M9
CANADA

TOTAL DIVING
IS IN COMPLIANCE WITH THE AIR/GAS QUALITY PORTION OF THE SPECIFICATION:
CSA STANDARD Z275.2-15 OCCUPATIONAL SAFETY CODE FOR DIVING OPERATIONS (H)
AS ANALYZED AND REPORTED ON THIS CERTIFICATE
FOR THE SAMPLE DESCRIBED UNDER SECTION "SAMPLE & REPORT INFORMATION"



American Assn for Laboratory Accreditation
1991: Certificate No. 322.01 Chemical Field of Testing

Richard A. Smith
Richard A. Smith, Laboratory Director

| Analytical Test Methods | Media Sampled | Estimate of Uncertainty |
|--|------------------------------|--|
| Gases & Vapors CAT-A-01 Gas Chromatography/Mass Spectrometry | Source Bottle: 737345 | The average analytical uncertainty (k=2) is 98.8±2.4% (relative) at the specification limit for the ten compounds normally reported. For uncertainty information for a specific compound, contact Trace Analytics. |
| Oil & Particulate CAT-A-03 Analytical Gravimetry | Ambient Bottle: 435033 | |
| Particle Size CAT-A-04 Optical Microscopy | Source Filter: 4137 | |
| Pressure Dew Point CAT-A-07 Gas Detector Tube | Detector Tube: Draeger 5-a/P | |

Results relate only to items tested. This report shall not be reproduced except in full without the written permission of Trace Analytics, LLC
© Copyright 2017, Trace Analytics, LLC

Sample & Report Information

Results of Test: PASS

| Sampled For | Total Diving | Analytes | Source Results | Ambient Results | Specification ¹ Allowable Limits |
|-------------------|-------------------------|---|----------------|-----------------|---|
| Sampled By | Not Provided | Oxygen, Volume % | 21.1 | 21.0 | 20-22 |
| Sampled On | 6/21/2017 | Nitrogen, Volume % | 78.0 | 78.1 | N/A |
| Received On | 6/26/2017 | Argon, Volume % | 0.9 | 0.9 | N/A |
| Analyzed On | 6/26/2017 | Nitrogen Plus Argon, Volume % | 78.9 | N/A | 78-80 |
| Sampled From | Compressor & Stored Air | Carbon Monoxide (CO), ppmv | <0.3 | <0.3 | 3 |
| Make | Jordair | Carbon Dioxide (CO ₂), ppmv | 348 | 373 | 600 |
| Model | K15-3EV | Water Content (H ₂ O), ppmv/Dewpoint, °F | <3.4 / <-91 | N/A | 27 / -63 (W) |
| Serial No. | 82/23/6/01 | Atmospheric Dew Point, °F (DT) | -86 | N/A | N/A |
| Cylinder(s) | 4 | TVHC (including CH ₄), ppmv | 4.7 | 5.4 | 15 |
| | | Methane (CH ₄) ppmv | 2.2 | 2.2 | 10 |
| | | TVHC (excluding CH ₄), ppmv | 2.5 | 3.2 | 5 |
| | | Oil (condensed) & Particulate, mg/m ³ | <0.03 | N/A | 1 |
| Hours | 4040 | Odor (provided by customer) | None/Slight | N/A | None/Slight |
| Sample Phase | Routine | Halogenated Hydrocarbons, ppmv | <0.1 | N/A | 5 |
| Customer Comments | | Atmospheric Dewpoint, °C | <-68 | N/A | -53 |
| | | Pressure Dewpoint, °C | <-53 | N/A | -5 |

PASS

(H) Compressed breathing air in cylinders and piping ≥ 15.3 MPa (2216 psig) shall have an atmospheric dew point ≤ -53°C (-63°F) or water vapor ≤ 27 mL/m³ (ppm) and SHOULD have a pressure dew point ≤ 5°C (9°F) below the lowest temperature to which the cylinder or piping can be exposed during any time of the year at that geographic location. If an operating pressure is not provided, we will use 20.7 MPag (3000 psig) with the Greenspan water vapour enhancement factor applied.
(W) Dew point is expressed in °F at one atmosphere pressure absolute.
(DT) Dew point is calculated at 1 atmosphere pressure (14.7 psia) from the detector tube reading.

Report Number 17-18938
Customer ID 2292
Date Reported 6/27/2017
Frequency Quarterly
Next Sample Due Approx. **9/21/2017**